



An ISO 9001 & ISO 13485 certified company  
**CERTIFICATE OF ANALYSIS**

Mfg. Lic. No. : 874/L

Date: 12/07/2018

**PRODUCT** : Truenat™ Malaria Pv/Pf  
**LOT. NO.** : ML005  
**MFG. DATE** : 2018- 06  
**EXP. DATE** : 2020- 05  
**PACK SIZE** : 5T, 20T  
**STORAGE** : 2°C to 30°C

**A.R. NO.** : B-ML-005

**PRECAUTION:** The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 nos, 20 nos.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01 no.	Pass

Performance				
Sr. No.	CONTROLS	Test Lot (Ct Values)		
Specificity Tests		Pv	Pf	IC
Negative Control		ND	ND	28.8
Positive Control		28.56	28.67	27.38
		28.4	29.2	28.11
		29	29.67	27
		28.29	30.38	26.75
Standard deviation		± 0.31	± 0.72	± 0.59

\*ND- Not detected

\*A deviation of ± 1 is acceptable for Ct values

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**

  
**QUALITY CONTROL**  
**Name: Dr. Praveen Kumar**  
**Designation: QC Head**

Page 1 of 1

**Molbio Diagnostics Pvt. Ltd.**

**Regd. Office:** H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA  
**Works:** Plot Nos. M 46-47, Phase III - B, Verna Industrial Estate, Verna - Goa 403 722. INDIA  
**Tel:** 91 832 2783267/6624783 **E mail:** admin@molbiodiagnostics.com  
**Email:** sales@molbiodiagnostics.com **Website:** www.molbiodiagnostics.com  
**CIN:** U33125GA2000PTC002909