



MOLBIO DIAGNOSTICS PVT. LTD.
An ISO 9001 & ISO 13485 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No. : 874/L

Date: 26/12/2018

PRODUCT : Trueprep™ AUTO Universal Cartridge Based Sample prep kit

Trueprep™ AUTO Universal
Cartridge Based
Sample prep kit

Lot No.: USPK084 Mfg Date: 2018-11 Exp Date: 2020-10

A.R. No.: B-USPK-084

Pack size : 5, 25 TESTS

STORAGE : 2°C to 40°C

PRECAUTION: The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton(Trueprep™ AUTO Universal Cartridge Based Sample prep kit [Present (P)/Not Present (NP)])		
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

Sr.No.	Sample Neagtive/Positive	Negative:TEST-ND	Control-Detected
		Positive:TEST-Detected	Control-Detected
		TEST	Control
01	Negative MTB	ND	29.13
02	Positive MTB	33.0	29.2
03	Negative HBV	ND	27.43
04	Positive HBV	30.33	28.5
05	Positive HCV	29.11	27.8
06	Positive HIV	30.0	27.0

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE



QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Molbio Diagnostics Pvt. Ltd.

Regd. Office: H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA
Works: Plot Nos. M 46-47, Phase III - B, Verna Industrial Estate, Verna - Goa 403 722. INDIA
Tel: 91 832 2783267/6624783 E mail: admin@molbiodiagnostics.com
Office: No. UGO-7, Ground Floor, Ambrosia Residency, Cabesa, Alto Santacruz, Goa - 403 005. INDIA
Tel: 91 832 2458111/2458112 Email: sales@molbiodiagnostics.com
Website: www.molbiodiagnostics.com CIN: U33125GA2000PTC002909