



MOLBIO DIAGNOSTICS PVT. LTD.  
An ISO 9001 & ISO 13485 certified company

### CERTIFICATE OF ANALYSIS

Mfg.Lic. No. : 874/L

Date: 03/01/2019

<b>PRODUCT : Trueprep™ AUTO Universal Cartridge Based Sample prep kit</b>	
<b>Trueprep™ AUTO Universal Cartridge Based Sample prep kit</b>	<b>Lot No.: USPK086 Mfg Date: 2018-11 Exp Date: 2020-10</b>
	<b>A.R. No.: B-USPK-086</b>
<b>Pack size : 5, 25 TESTS</b>	<b>STORAGE : 2°C to 40°C</b>

**PRECAUTION:** The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
<b>A</b>	<b>Contents in the carton(Trueprep™ AUTO Universal Cartridge Based Sample prep kit [Present (P)/Not Present (NP)])</b>		
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

Sr.No.	Sample Neagtive/Positive	Negative:TEST-ND Positive:TEST-Detected	Control-Detected Control-Detected
		TEST	Control
01	Negative MTB	ND	26.67
02	Positive MTB	29.8	27.67
03	Negative HBV	ND	29.33
04	Positive HBV	26.0	30.10
05	Negative HCV	ND	26.5
06	Positive HIV	33.6	28.33

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**

**QUALITY CONTROL**  
Name: Dr. Praveen Kumar  
Designation: QC Head

## Molbio Diagnostics Pvt. Ltd.

Regd. Office: H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA  
Works: Plot Nos. M 46-47, Phase III - B, Verna Industrial Estate, Verna - Goa 403 722. INDIA  
Tel: 91 832 2783267/6624783 E mail: admin@molbiodiagnostics.com  
Office: No. UGO-7, Ground Floor, Ambrosia Residency, Cabesa, Alto Santacruz, Goa - 403 005. INDIA  
Tel: 91 832 2458111/2458112 Email: sales@molbiodiagnostics.com  
Website: www.molbiodiagnostics.com: CIN: U33125GA2000PTC002909