



MOLBIO DIAGNOSTICS PVT. LTD.
An ISO 9001 & ISO 13485 certified company
CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : 874/L

Date: 21/09/2018

PRODUCT : Trueprep™ AUTO MTB Sample Pre-treatment Pack		
Lot No.: MSPT010	Mfg Date: 2018-06	Exp Date: 2020-05
A.R. No.: B-MSPT-010		
Pack size : 20 TESTS , 5 TESTS	STORAGE : 2°C to 40°C	

PRECAUTION: The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
D	Contents in the carton(MTB Sample Pre-treatment Pack)(Present (P)/Not Present (NP))		
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1ml)	Present	Present

Performance on True Lab Uno					
Sr.No.	CONTROLS	Specification (Ct Values)		Test Lot (Ct Values)	
Specificity Tests		TEST	IC	TEST	IC
Negative Control					
1	Sample 1	ND	IC should be detected	ND	28.14
Positive Control					
1	Sample 1	Test and IC should be detected		27.33	27.67
2	Sample 2	Test and IC should be detected		26.83	27.4
3	Sample 3	Test and IC should be detected		27.0	28.29

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Page 1 of 1

Molbio Diagnostics Pvt. Ltd.

Regd. Office: H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA

Works: Plot Nos. M 46-47, Phase III - B, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: 91 832 2783267/6624783 E mail: admin@molbioldiagnostics.com

Office: No. UGO-7, Ground Floor, Ambrosia Residency, Cabesa, Alto Santacruz, Goa - 403 005. INDIA

Tel: 91 832 2458111/2458112 Email: sales@molbioldiagnostics.com

Website: www.molbioldiagnostics.com: CIN: U33125GA2000PTC002909