



MOLBIO DIAGNOSTICS PVT. LTD.  
An ISO 9001 & ISO 13485 certified company

### CERTIFICATE OF ANALYSIS

Mfg.Lic. No. : MFG/IVD/2019/000003

Date: 17/05/2019

PRODUCT : Trueprep™ AUTO Universal Cartridge Based Sample prep kit	
Trueprep™ AUTO Universal Cartridge Based Sample prep kit	Lot No.: USPK156 Mfg Date: 2019-04 Exp Date: 2021-03
	A.R. No.: B-USPK-156
Pack size : 5, 25 TESTS	STORAGE : 2°C to 40°C

**PRECAUTION:** The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton(Trueprep™ AUTO Universal Cartridge Based Sample prep kit [Present (P)/Not Present (NP)])		
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

Sr.No.	Sample Negative/Positive	Negative:TEST-ND	Control-Detected
		Positive:TEST-Detected	Control-Detected
		TEST	Control
01	Negative MTB	ND	28.33
02	Positive MTB	29.4	28.0
03	Positive MTB	31.13	27.75
01	Negative MTB Plus	ND	28.33
02	Positive MTB Plus	29.4	29.6
03	Positive MTB Plus	30.11	26.8

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

*Fot*  
*Praveen Kumar*  
VERNA  
**QUALITY CONTROL**  
Name: Dr. Praveen Kumar  
Designation: QC Head

## Molbio Diagnostics Pvt. Ltd.

Regd. Office: H. No. 13, Sagar Society, Dona Paula, Panaji, North Goa 403 004. INDIA  
Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA  
Tel: +91-832-2783267 E mail: admin@molbiodiagnostics.com  
Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com  
CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO