



MOLBIO DIAGNOSTICS PVT. LTD.
An ISO 9001:2015 & ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 17/02/2020

Product: Trueprep [®] AUTO Universal Cartridge Based Sample prep kit	
Trueprep [®] AUTO Universal Cartridge Based Sample prep kit	Lot No.: USP271 Mfg Date: 2020-01 Exp Date: 2021-12
	A.R. No.: B-USPK-271
Pack size: 5, 25 TESTS	Storage: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep [®] AUTO Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab [™] micro PCR Analyzers			
Sr. No.	Sample Negative/Positive	Negative: Test- ND	IC- Detected
		Positive: Test- Detected	IC- Detected
		Test (Ct)	IC (Ct)
1	Negative MTB	ND	28.17
2	Positive MTB	31.83	29.8
3	Negative HBV	ND	27.5
4	Positive HBV	30.29	29.8
5	Positive HCV	26.55	29.5
6	Positive HIV	27.8	26.6

*Ct- Cycle threshold

*IC- Internal Control

*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE



Name: Dr. Rakesh Kumar
Designation: QC Head

Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO