



MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 25/01/2021

BRAND NAME	Truenat™ Positive Control Kit - Panel I		
GENERIC NAME	Control Kit for Truenat™ Chip-based micro PCR Tests: MTB, MTB Plus, MTB-RIF Dx, SARS CoV-2, COVID-19 and H1N1		
LOT NO	PCI002		
MFG. DATE	2021-01		
EXP. DATE	2021-12		
PACK SIZE	08T		
STORAGE	2°C to 40°C		A.R. No. : B-PCI-002


PRECAUTION: Refer Safety Data Sheet and instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Negative control	Present	Present
2	Positive control (08- strip tubes containing dried down positive control)	Present	Present
3	Package insert	Present	Present

Specification (Ct Value)				Observation (Ct Value)	
Sr. No.	Truenat™ test name	Positive control	Negative control	Positive control	Negative control
1	MTB	28.1 ± 2	ND	28.1	ND
2	MTB Plus	27.4 ± 2	ND	27.4	ND
3	MTB-RIF Dx	RIF Resistance not detected	Indeterminate	RIF Resistance not detected	Indeterminate
4	SARS CoV-2	27.8 ± 2	ND	27.8	ND
5	COVID-19	ORF1A: 27.0 ± 2; E gene: 29.0 ± 2	ND	ORF1A: 27.0 E gene: 29.0	ND
6	H1N1	29.0 ± 2	ND	29.0	ND

*ND- Not detected, Ct- Cycle threshold, RIF- Rifampicin

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head
Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO