



MOLBIO DIAGNOSTICS PRIVATE LIMITED  
ISO 13485:2016 certified company

## CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 26/01/2021

Product: Trueprep <sup>®</sup> AUTO v2 Universal Cartridge Based Sample Prep Kit	
Trueprep <sup>®</sup> AUTO v2 Universal Cartridge Based Sample Prep Kit	Lot No.: VUCPK081 Mfg Date: 2021-01 Exp Date: 2022-12
	A.R. No.: B-VUCPK-081
Pack size: 50 TESTS	Storage: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep <sup>®</sup> AUTO v2 Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab <sup>®</sup> micro PCR Analyzers			
Sr. No.	Sample Negative/Positive	Negative: Test- ND	IC- Detected
		Positive: Test- Detected	IC- Detected
		Test (Ct)	IC (Ct)
1	Negative MTB	ND	27.17
2	Negative MTB	ND	27.8
3	Negative MTB	ND	27.33
4	Positive MTB	29.6	27.17
5	Positive MTB	29.33	27.13
6	Positive MTB	29.67	27.0

\*Ct- Cycle threshold

\*IC- Internal Control

\*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

  
QUALITY CONTROL  
Name: Dr. Praveen Kumar  
Designation: QC Head

Page 1 of 1

### Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1Z0