



MOLBIO DIAGNOSTICS PVT. LTD.  
An ISO 9001:2015 & ISO 13485:2016 certified company

### CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 22/01/2020

|   |  |
|---|--|
| Product: Trueprep <sup>®</sup> AUTO Universal Cartridge Based Sample prep kit |  |
| Trueprep <sup>®</sup> AUTO Universal Cartridge Based Sample prep kit          | Lot No.: USPK257 Mfg Date: 2019-10 Exp Date: 2021-09 |
|   | A.R. No.: B-USPK-257                                 |
| Pack size: 5, 25 TESTS  | Storage: 2°C to 40°C                                 |

**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

| Sr. No. | Components in each kit  | Specification                  | Observation |
|---------|---|--------------------------------|-------------|
| A       | Contents in the carton (Trueprep <sup>®</sup> AUTO Universal Cartridge Based Sample prep kit) | (Present (P)/Not Present (NP)) |             |
| 1       | Universal Reagent Pack  | Present                        | Present     |
| 2       | Universal Cartridge Pack  | Present                        | Present     |
| 3       | Disposable Transfer Pipette (3 ml)  | Present                        | Present     |
| 4       | Package Insert  | Present                        | Present     |

| Performance on Truelab <sup>™</sup> micro PCR Analyzers |                          |                          |              |
|---|--------------------------|--------------------------|--------------|
| Sr. No.   | Sample Negative/Positive | Negative: Test- ND       | IC- Detected |
|   |                          | Positive: Test- Detected | IC- Detected |
|   |                          | Test (Ct)                | IC (Ct)      |
| 1   | Negative HCV             | ND                       | 30.0         |
| 2   | Negative HCV             | ND                       | 30.57        |
| 3   | Positive HCV             | 28.78                    | 28.63        |
| 4   | Positive HCV             | 29.7                     | 30.14        |
| 5   | Positive HCV             | 28.07                    | 28.2         |
| 6   | Positive HCV             | 29.45                    | 30.29        |

\*Ct- Cycle threshold

\*IC- Internal Control

\*ND- Not Detected

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**



**QUALITY CONTROL**  
Name: Dr. Praveen Kumar  
Designation: QC Head

## Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO