



MOLBIO DIAGNOSTICS PRIVATE LIMITED
ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 23/02/2021

Product: Trueprep [®] AUTO v2 Universal Cartridge Based Sample Prep Kit	
Trueprep [®] AUTO v2 Universal Cartridge Based Sample Prep Kit	Lot No.: VUCPK090-P Mfg Date: 2021-01 Exp Date: 2022-12
	A.R. No.: B-VUCPK-090-P
Pack size: 50 TESTS	Storage: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep [®] AUTO v2 Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present


Performance on Truelab [®] micro PCR Analyzers			
Sr. No.	Sample Negative/Positive	Negative: Test- ND	IC- Detected
		Positive: Test- Detected	IC- Detected
		Test (Ct)	IC (Ct)
1	Negative MTB	ND	27.33
2	Positive MTB	27.6	28.33
3	Negative HBV	ND	26.0
4	Positive HBV	29.88	26.75
5	Positive HCV	29.5	30.33
6	Negative HIV	ND	28.75

*Ct- Cycle threshold

*IC- Internal Control

*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE



QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO