

## MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

## **CERTIFICATE OF ANALYSIS**

Mfg.Lic. No.: MFG/IVD/2020/000119

Date: 16/02/2021

PRODUCT: Trueprep® AUTO Universal Sample Pre-treatment Pack

Trueprep® AUTO Universal Lot No.: USPT142 Mfg Date: 2021-01 Exp Date: 2022-12

Sample Pre-treatment Pack A.R. No.: B-USPT-142

Pack size: 50T, 20T, 5T STORAGE: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
A	Contents in the carton (Trueprep® AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Disposable Transfer Pipette (1ml)	Present	Present	

	Perform	nance on	Truelab <sup>®</sup> micro PCR Analyzers	3	
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negat	ive Control				
1	Neg HBV	ND	Control should be detected	ND	30.25
2	Neg HIV	ND	Control should be detected	ND	29.8
3	Neg COVID-19	ND	Control should be detected	ND	26.0
Positi	ve Control		*		Marie
1	Pos HBV	Test and Control should be detected		29.71	28.18
2	Pos HCV	Test and Control should be detected		27.78	29.67
3	Pos COVID-19	Test an	d Control should be detected	27.8	24.25

<sup>\*</sup>Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE.

QUALITY CONTROL Name: Dr. Praveen Kumar Designation:QC Head

Page 1 of 1

## **Molbio Diagnostics Private Limited**

<sup>\*</sup>IC- Internal Control

<sup>\*</sup>ND- Not Detected