

MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No. : MFG/IVD/2020/000003

Date: 10/02/2021

PRODUCT: Trueprep® AUTO Universal Sample Pre-treatment Pack

Trueprep® AUTO Universal Sample Pre-treatment Pack

Lot No.: USPT144 Mfg Date: 2021-01 Exp Date: 2022-12

A.R. No.: B-USPT-144

Pack size: 50T, 20T, 5T

STORAGE: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

| Sr. No. | Components in each kit | Specification | Observation | |
|---------|--|---------------------------------|-------------|--|
| Α | Contents in the carton (Trueprep® AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP)) | (Present (P)/ Not Present (NP)) | | |
| 1 | Lyoio Duffer | Present | Present | |
| 2 | Disposable Transfer Pipette (1ml) | Present | Present | |

| | Perfor | mance on | Truelab® micro PCR Analyzers | | 74 |
|------------------|-------------------|-------------------------------------|------------------------------|-------------------------|------|
| Sr. No. | Control | Specification (Ct Values) | | Test Lot (Ct Values) | |
| 1 | Specificity Tests | Test | IC | Test | 10 |
| Negative Control | | | | | IC |
| 1 | Neg HBV | ND | Control should be detected | ND | 28.8 |
| 2 | Neg HIV | ND | Control should be detected | ND | 28 |
| 3 | Neg COVID-19 | ND | Control should be detected | ND | 24 |
| Positiv | e Control | | | | |
| 1 | Pos HBV | Test and Control should be detected | | | |
| 2 | Pos HCV | Test and Control should be detected | | 31.14 | 30 |
| 3 | Pos COVID-19 | Test and Control should be detected | | 27.7 | 28 |
| Ct- Cycle | e threshold | Test and Control should be detected | | 29.67 | 26.4 |

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS NOS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

> QUALITY CONTROL Name: Dr. Praveen Kumar Designation:QC Head

Page 1 of 1 Molbio Diagnostics Private Limited

^{*}IC- Internal Control

^{*}ND- Not Detected