

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2020/000003

Date: 11/02/2021

PRODUCT: Trueprep® AUTO Universal Sample Pre-treatment Pack

Trueprep® AUTO Universal Lot No.: US

Lot No.: USPT145 Mfg Date: 2021-02 Exp Date: 2023-01

Sample Pre-treatment Pack A.R. No.: B-USPT-145

Pack size: 50T, 20T, 5T STORAGE: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
Α	Contents in the carton (Trueprep® AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Disposable Transfer Pipette (1ml)	Present	Present	

	Perform	mance on	Truelab [®] micro PCR Analyzers	7-	
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negati	ve Control				
1	Neg HBV	ND	Control should be detected	ND	28.2
2	Neg HIV	ND	Control should be detected	ND	28.5
3	Neg COVID-19	ND	Control should be detected	ND	24.5
Positiv	re Control				
1	Pos HBV	Test and Control should be detected		29.6	28.5
2	Pos HCV	Test and Control should be detected		29.6	28.2
3	Pos COVID-19	Test and Control should be detected		27.5	22.33

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL Name: Dr. Praveen Kumar Designation:QC Head

Molbio Diagnostics Private Limited

^{*}IC- Internal Control

^{*}ND- Not Detected