

MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 02/03/2021

Product: Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit

Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit Lot No.: VUCPK097 Mfg Date: 2021-02 Exp Date: 2023-01

A.R. No.: B-VUCPK-097

Pack size: 50 TESTS

Storage: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

| Sr. No. | Components in each kit | S7ecification | Observation |
|---------|---|--------------------------------|-------------|
| Α | Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit) | (Present (P)/Not Present (NP)) | |
| 1 | Universal Reagent Pack | Present | Present |
| 2 | Universal Cartridge Pack | Present | Present |
| 3 | Disposable Transfer Pipette (3 ml) | Present | Present |
| 4 | Package Insert | Present | Present |

| Performance on Truelab® micro PCR Analyzers | | | | | | |
|---|-----------------------------|---|------------------------------|--|--|--|
| Sr. No. | Sample Negative/Positive | Negative: Test- ND Positive: Test- Detected | IC- Detected IC- Detected | | | |
| | | Test (Ct) | IC (Ct) | | | |
| 1 | Negative MTB | ND | 27.5 | | | |
| 2 | Negative MTB | ND | 28.0 | | | |
| 3 | Negative MTB | ND | 27.2 | | | |
| 4 | Positive MTB | 27.0 | 27.67 | | | |
| 5 | Positive MTB | 26.8 | 26.43 | | | |
| 6 | Positive MTB | 26.6 | 26.67 | | | |

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED

PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL Name: Dr. Praveen Kumar Designation: QC Head

Molbio Diagnostics Private Limited

^{*}IC- Internal Control

^{*}ND- Not Detected