



MOLBIO DIAGNOSTICS PVT. LTD.

An ISO 9001 & ISO 13485 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : 874/L

Date: 28-03-2018

PRODUCT : Truenat™ Dengue
LOT. NO. : DN013
MFG. DATE : 2018- 03
EXP. DATE : 2020- 02
PACK SIZE : 05T, 20T
STORAGE : 2°C to 30°C

A.R. NO. : B-DN-013

PRECAUTION: The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contain 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	05 nos, 20 nos.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01 no.	Pass

Performance			
Sr. No.	CONTROLS	Test Lot (Ct Values)	
Specificity Tests		DEN	IC
Negative Control		ND	27
Positive Control		26.67	28.29
		26	27.71
		26.5	27.25
		27.3	27.67
Standard deviation		± 0.53	± 0.42

*ND- Not detected

*A deviation of ± 1 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Page 1 of 1

Molbio Diagnostics Pvt. Ltd.

Regd. Office: H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA
Works: Plot Nos. M 46-47, Phase III - B, Verna Industrial Estate, Verna - Goa 403 722. INDIA
Tel: 91 832 2783267/6624783 E mail: admin@molbiodiagnostics.com
Office: No. UGO-7, Ground Floor, Ambrosia Residency, Cabesa, Alto Santacruz, Goa - 403 005. INDIA
Tel: 91 832 2458111/2458112 Email: sales@molbiodiagnostics.com
Website: www.molbiodiagnostics.com: CIN: U33125GA2000PTC002909