

MOLBIO DIAGNOSTICS PVT. LTD. An ISO 9001:2015 & ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 01/11/2019

PRODUCT : Trueprep™ AUTO MTB Sample Pre-treatment Pack

Trueprep™ AUTO MTB Lot No.: MSPT044 Mfg Date: 2019-09 Exp Date: 2021-08

Sample Pre-treatment Pack A.R. No.: B-MSPT-044

A.N. No.: D-WSF 1-044

Pack size: 20 TESTS, 5 TESTS STORAGE: 2°C to 40°C PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
Α	Contents in the carton(MTB Sample Pre-treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Liquefaction Buffer	Present	Present	
3	Disposable Transfer Pipette (1ml)	Present	Present	

Sr. No.	CONTROLS Specificity Tests	nce on Truelab™ Real Time micro PCR A Specification (Ct Values)		Test Lot (Ct Values)	
1		Test	IC	Test	IC
Neg	ative Control				
1	Sample 1	ND	Control should be detected	ND	27.8
Posi	tive Control				
1	Sample 1	Test a	nd Control should be detected	27.8	27.5
2	Sample 2	Test and Control should be detected		28.57	28
3	Sample 3	Test a	nd Control should be detected	29.13	28.33

*Ct- Cycle threshold *IC- Internal Control *ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS

PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL
Name: Dr. Prayeen Kumar
Designation:QC Head

Page 1 of 1