



MOLBIO DIAGNOSTICS PVT. LTD.  
An ISO 9001:2015 & ISO 13485:2016 certified company

**CERTIFICATE OF ANALYSIS**

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 10/12/2019

<b>PRODUCT: Trueprep™ AUTO MTB Sample Pre-treatment Pack</b>	
Trueprep™ AUTO MTB Sample Pre-treatment Pack	Lot No.: MSPT045 Mfg Date: 2019-09 Exp Date: 2021-08
	A.R. No.: B-MSPT-045
Pack size: 20 TESTS, 5 TESTS	STORAGE: 2°C to 40°C

**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1ml)	Present	Present


Performance on Truelab™ Real Time micro PCR Analyzers					
Sr. No.	CONTROLS	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
<b>Negative Control</b>					
1	Sample 1	ND	Control should be detected	ND	28.0
<b>Positive Control</b>					
1	Sample 1	Test and Control should be detected		28	27.2
2	Sample 2	Test and Control should be detected		28.14	27.13
3	Sample 3	Test and Control should be detected		28.25	27.67

\*Ct- Cycle threshold

\*IC- Internal Control

\*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE



QUALITY CONTROL  
Name: Dr. Praveen Kumar  
Designation: QC Head

**Molbio Diagnostics Pvt. Ltd.**

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA  
Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com  
Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com  
CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1Z0