



MOLBIO DIAGNOSTICS PVT. LTD.
An ISO 9001:2015 & ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : MFG/IVD/2019/000003

Date: 27/08/2019

PRODUCT : Trueprep™ AUTO MTB Sample Pre-treatment Pack	
Trueprep™ AUTO MTB Sample Pre-treatment Pack	Lot No.: MSPT038 Mfg Date: 2019-08 Exp Date: 2021-07
	A.R. No.: B-MSPT-038
Pack size : 20 TESTS, 5 TESTS	STORAGE : 2°C to 40°C
PRECAUTION: Refer Safety Data Sheet and follow instructions for use.	

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton(MTB Sample Pre-treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1ml)	Present	Present

Performance on Truelab™ Real Time micro PCR Analyzers					
Sr. No.	CONTROLS	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negative Control					
1	Sample 1	ND	Control should be detected	ND	29.75
Positive Control					
1	Sample 1	Test and Control should be detected		28	28
2	Sample 2	Test and Control should be detected		28	29.2
3	Sample 3	Test and Control should be detected		27.57	27.2

*Ct- Cycle threshold *IC- Internal Control *ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbioldiagnostics.com

Email: sales@molbioldiagnostics.com Website: www.molbioldiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO