



MOLBIO DIAGNOSTICS PVT. LTD.  
An ISO 9001 & ISO 13485 certified company  
**CERTIFICATE OF ANALYSIS**

Mfg. Lic. No. : MFG/IVD/2019/000003

Date: 02/07/2019

<b>PRODUCT : Trueprep™ AUTO MTB Sample Pre-treatment Pack</b>		
<b>Lot No.: MSPT031</b>	<b>Mfg Date: 2019-06</b>	<b>Exp Date: 2021-05</b>
<b>A.R. No.: B-MSPT-031</b>		
<b>Pack size : 20 TESTS , 5 TESTS</b>	<b>STORAGE : 2°C to 40°C</b>	

**PRECAUTION:** The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
D	Contents in the carton(MTB Sample Pre-treatment Pack)(Present (P)/Not Present (NP))		
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1ml)	Present	Present

Performance on Truelab™ Real Time micro PCR Analyser					
Sr.No.	CONTROL S	Specification (Ct Values)		Test Lot (Ct Values)	
Specificity Tests		TEST	Control	TEST	Control
<b>Negative Control</b>					
1	Sample 1	ND	Control should be detected	ND	29.2
<b>Positive Control</b>					
1	Sample 1	Test and Control should be detected		31.4	27.0
2	Sample 2	Test and Control should be detected		31.17	29.17
3	Sample 3	Test and Control should be detected		30.33	28.17

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**

  
**QUALITY CONTROL**  
Name: Dr. Praveen Kumar  
Designation: QC Head

Page 1 of 1

## Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1Z0