

## MOLBIO DIAGNOSTICS PVT. LTD. An ISO 9001:2015 & ISO 13485:2016 certified company

## CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 15/01/2020

PRODUCT : Trueprep™ AUTO Universal Sample Pre-treatment Pack

Trueprep™ AUTO Universal Lot No.: USPT030 Mfg Date: 2019-12 Exp Date: 2021-11

Sample Pre-treatment Pack A.R. No.: B-USPT-030

STORAGE: 2°C to 40°C Pack size: 20 TESTS, 5 TESTS

**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
A	Contents in the carton (Trueprep™ AUTO Universal Sample Pre-treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Disposable Transfer Pipette (1ml)	Present	Present	

	Perform	ance on	Truelab™ micro PCR Analyzer	S	
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negat	tive Control				
1	Neg HBV	ND	Control should be detected	ND	28.2
2	Neg HCV			ND	30.25
Positi	ive Control	-20118-20118-2011			
1	Pos HBV	Test and Control should be detected		31.0	29.33
2	Pos HIV	Test ar	nd Control should be detected	28.29	29.25

<sup>\*</sup>Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

> QUALITY CONTROL Name: Dr. Prayeen Kumar Designation: QC Head

Page 1 of 1

## Molbio Diagnostics Pvt. Ltd.

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO

<sup>\*</sup>IC- Internal Control

<sup>\*</sup>ND- Not Detected