

MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 13/04/2021

Doc ID: COA/TP/USPT159

PRODUCT: Trueprep® AUTO Universal Sample Pre-treatment Pack

Trueprep® AUTO Universal

Lot No.: USPT159 Mfg Date: 2021-04 Exp Date: 2023-03

Sample Pre-treatment Pack

A.R. No.: B-USPT-159

Pack size : 50T, 20T, 5T

STORAGE: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
A	Contents in the carton (Trueprep® AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Disposable Transfer Pipette (1ml)	Present	Present	

	Perform	mance on	Truelab [®] micro PCR Analyzers	;	
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negat	ive Control				
1	Neg HBV	ND	Control should be detected	ND	30.33
2	Neg COVID-19	ND	Control should be detected	ND	20.17
Positi	ve Control	***************************************			L
1	Pos HBV	Test and Control should be detected		29.8	27.5
2	Pos HCV	Test and Control should be detected		28.63	28.6
3	Pos HIV	Test and Control should be detected 28.5 29.6			
4	Pos COVID-19	Test and Control should be detected 29.11 25.17			

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL Name: Dr. Praveen Kumar Designation:QC Head

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO

^{*}IC- Internal Control

^{*}ND- Not Detected