

MOLBIO DIAGNOSTICS PVT. LTD. An ISO 9001:2015 & ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 10/12/2019

PRODUCT: Trueprep™ AUTO Universal Sample Pre-treatment Pack

Trueprep™ AUTO Universal Lot No.: USPT026 Mfg Date: 2019-11 Exp Date: 2021-10

Sample Pre-treatment Pack A.R. No.: B-USPT-026

Pack size: 20 TESTS, 5 TESTS

STORAGE: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

| Sr. No. | Components in each kit | Specification | Observation | |
|---------|---|---------------------------------|-------------|--|
| A | Contents in the carton (Trueprep™ AUTO Universal Sample Pre-treatment Pack)(Present (P)/Not Present (NP)) | (Present (P)/ Not Present (NP)) | | |
| 1 | Lysis Buffer | Present | Present | |
| 2 | Disposable Transfer Pipette (1ml) | Present | Present | |

| | Perform | nance on | Truelab™ micro PCR Analyzer | S | |
|------------|-------------------|-------------------------------------|-------------------------------|-------------------------|-------|
| Sr. No. | Control | Specification (Ct Values) | | Test Lot (Ct Values) | |
| 1 | Specificity Tests | Test | IC | Test | IC |
| Negat | tive Control | | | | |
| 1 | Neg HBV | ND | Control should be detected | ND | 29 |
| Positi | ive Control | | | | L |
| 1 | Pos HBV | Test and Control should be detected | | 29.75 | 30.14 |
| 2 | Pos HIV | Test and Control should be detected | | 28 | 27 |
| 3 | Pos HCV | Test ar | nd Control should be detected | 29.14 | 28.75 |

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACT⊕R PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

> QUALITY CONTROL Name: Dr. Praveen Kumar Designation:QC Head

Page 1 of 1

Molbio Diagnostics Pvt. Ltd.

^{*}IC- Internal Control

^{*}ND- Not Detected