



MOLBIO DIAGNOSTICS PRIVATE LIMITED
ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 25/11/2020

PRODUCT: Trueprep® AUTO MTB Sample Pre-treatment Pack	
Trueprep® AUTO MTB Sample Pre-treatment Pack	Lot No.: MSPT066 Mfg Date: 2020-10 Exp Date: 2022-09
	A.R. No.: B-MSPT-066
Pack size: 20 TESTS, 5 TESTS, 50 TESTS	STORAGE: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1ml)	Present	Present

Performance on Truelab® Real Time micro PCR Analyzers					
Sr. No.	Controls	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negative Control					
1	Sample 1	ND	Control should be detected	ND	27.4
Positive Control					
1	Sample 1	Test and Control should be detected		28.0	29.8
2	Sample 2	Test and Control should be detected		29.75	28.2
3	Sample 3	Test and Control should be detected		28.0	29.0

*Ct- Cycle threshold

*IC- Internal Control

*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO