



**MOLBIO DIAGNOSTICS PRIVATE LIMITED**

ISO 13485:2016 certified company

**CERTIFICATE OF ANALYSIS**

Mfg. Lic. No.: MFG/IVD/2019/000018

Date: 24/04/2021

Doc ID: COA/TN/CD074

<b>BRAND NAME</b>	Truenat™ COVID-19
<b>GENERIC NAME</b>	Chip-based Real Time Duplex PCR Test for COVID-19
<b>LOT No.</b>	CD074
<b>MFG. DATE</b>	2021-04
<b>EXP. DATE</b>	2023-03
<b>PACK SIZE</b>	5T, 20T, 50T
<b>STORAGE</b>	2°C to 30°C
<b>A.R. No.: B-CD-074</b>	

**PRECAUTION:** Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 dessicant pouch and 1 filter barrier tip	5 no, 20 no, 50 no.	Pass
2	Colour of desiccant pouch	Orange	Pass
3	Package insert	01 no.	Pass

PERFORMANCE			
CONTROLS	(Ct Values)		
Specificity Tests	Orf1A	Egene	Control
Negative Control	ND	ND	26
	ND	ND	25.43
Positive Control	25.5	25.14	26
	26.13	26.11	26.14
	26.14	26.67	26.14
	25.67	25.8	25.33
	25.67	26.29	25.75
	25.11	25.6	26
	25.38	25.67	25.25
	25.83	25.67	26
Standard deviation	+0.35	+0.47	+0.35

\*ND- Not detected, Ct- Cycle threshold

\*A deviation of  $\leq 1.5$  is acceptable for Ct values


**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS QC APPROVED FOR SALE**

**QUALITY CONTROL**

Name: Dr. Praveen Kumar

Designation: QC Head

Page 1 of 1



**Molbio Diagnostics Private Limited**

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1Z0