

MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003 Date: 05/05/2021

Doc ID: COA/TP/USPT180

PRODUCT: Trueprep® AUTO Universal Sample Pre-treatment Pack

Trueprep® AUTO Universal Lot No.: USPT180 Mfg Date: 2021-04 Exp Date: 2023-03

Sample Pre-treatment Pack A.R. No.: B-USPT-180

Pack size: 50T, 20T, 5T STORAGE: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use

Sr. No.	Components in each kit	Specification	Observation	
A	Contents in the carton (Trueprep® AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Disposable Transfer Pipette (1ml)	Present	Present	

	Perform	nance on	Truelab [®] micro PCR Analyzers	;	
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negat	ive Control				
1	Neg HBV	ND	Control should be detected	ND	29.6
2	Neg HCV	ND	Control should be detected	ND	28.5
3	Neg HIV	ND	Control should be detected	ND	28.33
4	Neg COVID-19	ND	Control should be detected	ND	25.14
5	Neg COVID-19	ND	Control should be detected	ND	24.6
Positi	ve Control				
6	Pos HBV	Test and Control should be detected		30.0	29.5

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL!
Name: Dr. Praveen Kumar
for Designation: QC Head

Page 1 of 1

^{*}IC- Internal Control

^{*}ND- Not Detected