

MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2020/000119

Date: 10/05/2021

Doc ID: COA/TP/USPT169

PRODUCT: Trueprep® AUTO Universal Sample Pre-treatment Pack

Trueprep® AUTO Universal Sample Pre-treatment Pack

Lot No.: USPT169 Mfg Date: 2021-04 Exp Date: 2023-03

A.R. No.: B-USPT-169

Pack size : 50T, 20T, 5T

STORAGE: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
A	Contents in the carton (Trueprep® AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Disposable Transfer Pipette (1ml)	Present	Present	

	Perforn	nance on	Truelab [®] micro PCR Analyzers	i	
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negat	ive Control				
1	Neg COVID-19	ND	Control should be detected	ND	20.0
2	Neg HBV	ND	Control should be detected	ND	28.2
Positi	ve Control		***************************************		
1	Pos HCV	Test and Control should be detected		29.82	29.0
2	Pos HBV	Test and Control should be detected		29.83	29.43
3	Pos HIV	Test an	d Control should be detected	29.75	27.8
4	Pos COVID-19	Test and Control should be detected 26.33 23.75			

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL Name: Dr. Praveen Kumar Designation:QC Head

Page 1 of 1

^{*}IC- Internal Control

^{*}ND- Not Detected