

MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 08/05/2021

Doc ID: COA/TP/USPT184

PRODUCT: Trueprep® AUTO Universal Sample Pre-treatment Pack

Trueprep® AUTO Universal Lot No.: USPT184 Mfg Date: 2021-04 Exp Date: 2023-03

Sample Pre-treatment Pack A.R. No.: B-USPT-184

Pack size: 50T, 20T, 5T STORAGE: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
Α	Contents in the carton (Trueprep® AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Disposable Transfer Pipette (1ml)	Present	Present	

	Perforr	nance on	Truelab® micro PCR Analyzers	3	
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negat	ive Control				<u> </u>
1	Neg HCV	ND	Control should be detected	ND	30.8
2	Neg COVID-19	ND	Control should be detected	ND	21.6
3	Neg Dengue	ND	Control should be detected	ND	29.8
Positi	ve Control				1
1	Pos HCV	Test and Control should be detected		29.88	28.5
2	Pos HIV	Test and Control should be detected		29.5	30.17
3	Pos COVID-19	Test an	nd Control should be detected	26.83	23.5

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL Name: Dr. Praveen Kumar Designation:QC Head

Page 1 of 1

Molbio Diagnostics Private Limited

^{*}IC- Internal Control

^{*}ND- Not Detected