



MOLBIO DIAGNOSTICS PRIVATE LIMITED  
ISO 13485:2016 certified company  
**CERTIFICATE OF ANALYSIS**

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 09/05/2021

Doc ID: COA/TP/VUCPK122-P

<b>Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit</b>	
<b>Trueprep® AUTO Universal Cartridge Based Sample Prep Kit</b>	<b>Lot No.: VUCPK122-P Mfg Date: 2021-03 Exp Date: 2023-02</b>
	<b>A.R. No.: VUCPK122-P</b>
<b>Pack size: 5, 25, 50 TESTS</b>	<b>Storage: 2°C to 40°C</b>

**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
<b>A</b>	<b>Contents in the carton (Trueprep® AUTO Universal Cartridge Based Sample prep kit)</b>	<b>(Present (P)/Not Present (NP))</b>	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

**Performance on Truelab® micro PCR Analyzers**

Sr. No.	Sample Negative/Positive	Negative: Test- ND	IC- Detected
		Positive: Test- Detected	IC- Detected
		Test (Ct)	IC (Ct)
1	Negative MTB	ND	29.43
2	Positive MTB	27.2	28.8
3	Positive MTB	27.13	29.0
4	Negative Dengue	ND	30.67
5	Positive HCV	31.27	30.5
6	Positive HCV	31.09	29.67

\*Ct- Cycle threshold

\*IC- Internal Control

\*ND- Not Detected

**Molbio Diagnostics Private Limited**

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO



MOLBIO DIAGNOSTICS PRIVATE LIMITED  
ISO 13485:2016 certified company  
**CERTIFICATE OF ANALYSIS**

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 09/05/2021

Doc ID: COA/TP/VUCPK122-P

Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit	
Trueprep® AUTO Universal Cartridge Based Sample Prep Kit	Lot No.: VUCPK122-P Mfg Date: 2021-03 Exp Date: 2023-02
	A.R. No.: VUCPK122-P
Pack size: 5, 25, 50 TESTS	Storage: 2°C to 40°C

Performance on Truelab® micro PCR Analyzers			
7	Positive HIV	30.17	29.2
8	Negative COVID-19	ND	22.6

\*Ct- Cycle threshold

\*IC- Internal Control

\*ND- Not Detected

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**



**QUALITY CONTROL**  
Name: Dr. Praveen Kumar  
Designation: QC Head

## Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO