

MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 28/04/2021

Doc ID: COA/TP/ATM059

Brand Name: Trueprep® AUTO Transport Medium for Swab Specimen Pack

Generic name: Medium for collection, Decontamination and Transport of Swab

Specimen for Trueprep® AUTO

Lot No.: ATM059 Mfg Date: 2021-04 Exp Date: 2023-03 A.R. No.: B-ATM-059

Pack size: 50T, 20T, 5T Storage Temperature: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep® AUTO Transport Medium for Swab Specimen Pack)	(Present (P)/ Not Present (NP))	
1	Transport Medium	Present	Present

	Performa	nce on Tr	uelab [®] micro PCR Ana	lyzers	
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Nega	ative Control				
1	Negative COVID-19	IC should be	ND	24.0	
2	Negative COVID-19		IC should be detected	ND	24.5
3	Negative COVID-19	ND		ND	22.0
Posi	tive Control				•
4	Positive COVID-19	Test and IC should be detected		28.43	25.5
5	Positive COVID-19	Test and IC should be detected		27.67	26.2
6	Positive COVID-19	Test and	IC should be detected	28.17	26.75

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE.

QUALITY CONTROL

Name: Dr. Praveen Kuma Designation: QC Head

Page 1 of 1

^{*}IC- Internal Control

^{*}ND- Not Detected