



MOLBIO DIAGNOSTICS PVT. LTD.  
An ISO 9001:2015 & ISO 13485:2016 certified company

### CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 02/03/2020

<b>Product: Trueprep<sup>®</sup> AUTO Universal Cartridge Based Sample prep kit</b>	
<b>Trueprep<sup>®</sup> AUTO Universal Cartridge Based Sample prep kit</b>	<b>Lot No.: USPK276 Mfg Date: 2020-01 Exp Date: 2021-12</b> <b>A.R. No.: B-USPK-276</b>
<b>Pack size: 5, 25 TESTS</b>	<b>Storage: 2°C to 40°C</b>

**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep <sup>®</sup> AUTO Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab <sup>™</sup> micro PCR Analyzers			
Sr. No.	Sample Negative/Positive	Negative: Test- ND	IC- Detected
		Positive: Test- Detected	IC- Detected
		Test (Ct)	IC (Ct)
1	Negative MTB	ND	30
2	Positive MTB	30.14	29.8
3	Positive MTB	30.75	31.0
4	Positive MTB	30.8	29.8

\*Ct- Cycle threshold

\*IC- Internal Control

\*ND- Not Detected

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**



Name: Dr. Praveen Kumar  
Designation: QC Head

## Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO