



MOLBIO DIAGNOSTICS PRIVATE LIMITED
ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No. : MFG/IVD/2019/000003

Date: 04/06/2021

Doc ID: COA/TP/USPT203

PRODUCT : Trueprep® AUTO Universal Sample Pre-treatment Pack	
Trueprep® AUTO Universal Sample Pre-treatment Pack	Lot No.: USPT203 Mfg Date: 2021-05 Exp Date: 2023-04
	A.R. No.: B-USPT-203
Pack size : 50T, 20T, 5T	STORAGE : 2°C to 40°C

.PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep® AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Disposable Transfer Pipette (1ml)	Present	Present

Performance on Truelab® micro PCR Analyzers					
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negative Control					
1	Neg HCV	ND	Control should be detected	ND	31.0
2	Neg COVID-19	ND	Control should be detected	ND	19.4
Positive Control					
1	Pos HIV	Test and Control should be detected		31.2	30.0
2	Pos HCV	Test and Control should be detected		31.0	29.5
3	Pos DENGUE	Test and Control should be detected		29.4	30.3
4	Pos DENGUE	Test and Control should be detected		28.89	29.45

*Ct- Cycle threshold

*IC- Internal Control

*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO