



MOLBIO DIAGNOSTICS PRIVATE LIMITED
ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2020/000073

Date: 06/06/2021

Doc ID: COA/TP/ATM062

Brand Name: Trueprep[®] AUTO Transport Medium for Swab Specimen Pack			
Generic name: Medium for collection, Decontamination and Transport of Swab Specimen for Trueprep[®] AUTO			
Lot No.: ATM062	Mfg Date: 2021-05	Exp Date: 2023-04	A.R. No.: B-ATM-062
Pack size: 50T, 20T, 5T		Storage Temperature: 2°C to 40°C	

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep [®] AUTO Transport Medium for Swab Specimen Pack)	(Present (P)/ Not Present (NP))	
1	Transport Medium	Present	Present

Performance on Truelab[®] micro PCR Analyzers

Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
		Test	IC	Test	IC
Negative Control					
1	Negative COVID-19	ND	IC should be detected	ND	26.6
2	Negative COVID-19			ND	23.29
3	Negative COVID-19			ND	23.5
Positive Control					
4	Positive COVID-19	Test and IC should be detected		25.2	25.0
5	Positive COVID-19	Test and IC should be detected		25.17	26.0
6	Positive COVID-19	Test and IC should be detected		26.0	25.0

*Ct- Cycle threshold

*IC- Internal Control

*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO