



MOLBIO DIAGNOSTICS PVT. LTD.

An ISO 9001 & ISO 13485 certified company

### CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : MFG/IVD/2019/000018

Date: 03/08/2019

PRODUCT : Truenat™ MTB

LOT. NO. : TB105

MFG. DATE : 2019- 07

EXP. DATE : 2021- 06

PACK SIZE : 5T, 20T

STORAGE : 2°C to 30°C

A.R. No. : B-TB-105b

**PRECAUTION:** The product is offered without warranty for any other specific use than that specified in the Instructions for Use.


Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 nos., 20 nos.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01 no.	Pass

Performance			
Sr. No.	CONTROLS	Test Lot (Ct Values)	
	Specificity Tests	MTB	Control
	Negative Control	ND	27.14
		ND	26.60
	Positive Control	29.0	26.50
		29.0	27.33
		29.40	28.17
		28.0	28.50
		28.14	26.56
		28.14	29.17
		28.40	27.60
		28.17	26.50
	Standard deviation	± 0.53	± 1.01

\*ND- Not detected

\*A deviation of  $\leq 1.5$  is acceptable for Ct values

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**

For  
  
QUALITY CONTROL  
Name: Dr. Praveen Kumar  
Designation: QC Head

Page 1 of 1

## Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnosics.com

Email: sales@molbiodiagnosics.com Website: www.molbiodiagnosics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO