

## MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

## **CERTIFICATE OF ANALYSIS**

Mfg.Lic. No.: MFG/IVD/2020/000119

Date: 29/06/2021

Doc ID: COA/TP/USPT214

PRODUCT: Trueprep® AUTO Universal Sample Pre-treatment Pack

Trueprep® AUTO Universal Sample Pre-treatment Pack

Lot No.: USPT214 Mfg Date: 2021-05 Exp Date: 2023-04

A.R. No.: B-USPT-214

Pack size : 50T, 20T, 5T

STORAGE: 2°C to 40°C

.PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
Α	Contents in the carton (Trueprep® AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Disposable Transfer Pipette (1ml)	Present	Present	

	Perform	mance on	Truelab <sup>®</sup> micro PCR Analyzers	<del></del>	
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negati	ive Control				1
1	Neg HBV	ND	Control should be detected	ND	30.6
2	Neg COVID-19	ND	Control should be detected	ND	22.17
Positiv	ve Control				
1	Pos HIV	Test and Control should be detected		30.8	28.8
2	Pos HCV	Test and Control should be detected		30.27	30.31
3	Pos HBV	Test and Control should be detected		29.2	30.33
4	Pos COVID-19	Test and Control should be detected 29.29			26.5

<sup>\*</sup>Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS

PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL Name: Dr. Praveen Kumar Designation: QC Head

Page 1 of 1

<sup>\*</sup>IC- Internal Control

<sup>\*</sup>ND- Not Detected