



MOLBIO DIAGNOSTICS PRIVATE LIMITED
ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No. : MFG/IVD/2019/000003

Date: 15/07/2021

Doc ID: COA/TP/USPT215

PRODUCT : Trueprep[®] AUTO Universal Sample Pre-treatment Pack	
Trueprep[®] AUTO Universal Sample Pre-treatment Pack	Lot No.: USPT215 Mfg Date: 2021-05 Exp Date: 2023-04
	A.R. No.: B-USPT-215
Pack size : 50T, 20T, 5T	STORAGE : 2°C to 40°C

.PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep[®] AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Disposable Transfer Pipette (1ml)	Present	Present


Performance on Truelab[®] micro PCR Analyzers


Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
		Test	IC	Test	IC
Negative Control					
1	Neg HBV	ND	Control should be detected	ND	28.17
Positive Control					
1	Pos HIV	Test and Control should be detected		31.0	30.0
2	Pos HIV	Test and Control should be detected		30.8	29.0
3	Pos HBV	Test and Control should be detected		28.0	29.8
4	Pos HBV	Test and Control should be detected		28.14	28
5	Pos COVID-19	Test and Control should be detected		28.33	27.6
6	Pos HCV	Test and Control should be detected		29.2	27.0
7	Pos HCV	Test and Control should be detected		29.23	30.13

*Ct- Cycle threshold

*IC- Internal Control

*ND- Not Detected
MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head



Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO