



MOLBIO DIAGNOSTICS PVT. LTD.

An ISO 9001 & ISO 13485 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : MFG/IVD/2019/000018

Date: 13-08-19

PRODUCT :Truenat HBV
LOT. NO. : HB039
MFG. DATE : 2019- 08
EXP. DATE : 2021- 07
PACK SIZE : 5 T, 20 T
STORAGE : 2°C to 30°C

A.R. NO. : B-HB-039

PRECAUTION: The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contain 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 nos. 20 nos.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01no.	Pass

Performance			
Sr. No.	CONTROLS	Test Lot (Ct Values)	
		HBV	Control
Specificity Tests			
Negative Control		ND	28.17
Positive Control		28.5	28.43
		27.4	28.0
		28.29	27.0
		29.5	27.57
Standard deviation		± 0.86	±0.61

*ND- Not detected

*A deviation of ≤ 1.5 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE


Name: Dr. Praveen Kumar
Designation: QC HEAD

Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO