

MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.:MFG/IVD/2019/000003

Date:25/09/2021

Doc.ID:COA/TP/USPT222

Product: Trueprep® AUTO Universal Sample Pre-treatment Pack					
Lot No.: USPT222	Mfg Date: 2021-06	Exp Date: 2023-05			
Pack size: 50T, 20T, 5T	Storage: 2°C to 40°C	A.R. No.: B-USPT-222			

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO Universal Sample Pre-treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Disposable Transfer Pipette (1ml)	Present	Present

	Performa	ance on	Truelab [®] micro PCR Analyzers		
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negati	ve Control				
1	Negative COVID-19	ND	Control should be detected	ND	26.8
2	Negative DENGUE	ND	Control should be detected	ND	29.13
3	Negative HCV	ND	Control should be detected	ND	28.8
Positiv	re Control				
1	Positive HIV	Test and Control should be detected		29.0	30
2	Positive HCV	Test and Control should be detected		29.6	29
3	Positive HCV	Test a	nd Control should be detected	27.56	29.33
4	Positive HBV	Test a	nd Control should be detected	27.17	29.33
5	Positive COVID-19	Test a	nd Control should be detected	30.29	24.17

*Ct- Cycle threshold

*IC- Internal Control

*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURING RECORDS ARE VERIFIED AND LOT IS RELEASED FOR SALE.

QUALITY CONTROL Name: Dr. Praveen Kumar Designation:QC Head

Page 1 of 1