

## MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

## **CERTIFICATE OF ANALYSIS**

Mfg. Lic. No.:MFG/IVD/2019/000003

Date:04/08/2021

Doc.ID:COA/TP/USPT220

Product: Trueprep® AUTO Universal Sample Pre-treatment Pack						
Lot No.: USPT220	Mfg Date: 2021-06	Exp Date: 2023-05				
Pack size: 50T, 20T, 5T	Storage: 2°C to 40°C	A.R. No.: B-USPT-220				

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
A	Contents in the carton (Trueprep® AUTO Universal Sample Pre-treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Disposable Transfer Pipette (1ml)	Present	Present	

14	Perform	ance on	Truelab <sup>®</sup> micro PCR Analyzers		
Sr. No.	Control Specification (Ct Values)		pecification (Ct Values)	Test Lot (Ct Values)	
1	Specificity Tests	Test	IC 4	Test	IC
Negat	ive Control				
1	Negative HBV	ND	Control should be detected	ND	28.4
2	Negative COVID-19	ND	Control should be detected	ND	25.17
3	Negative COVID-19	ND	Control should be detected	ND	26.5
Positi	ve Control				
1	Positive HIV	Test and Control should be detected		28.6	29.4
2	Positive HIV	Test and Control should be detected		28.5	30.6
3	Positive HBV	Test ar	nd Control should be detected	27	29.8
4	Positive HBV	Test ar	nd Control should be detected	27.29	30.6
5	Positive COVID-19	Test ar	nd Control should be detected	29.14	24
6	Positive HCV	Test ar	nd Control should be detected	30.6	30

<sup>\*</sup>Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation:QC Head

Page 1 of 1

## **Molbio Diagnostics Private Limited**

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO

<sup>\*</sup>IC- Internal Control

<sup>\*</sup>ND- Not Detected