



MOLBIO DIAGNOSTICS PVT. LTD.

An ISO 9001:2015 & ISO 13485:2016 Certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : MFG/IVD/2019/000040

Date: 13-12-2019

BRAND NAME	Truenat™ HCV
GENERIC NAME	Chip-based Real Time PCR Test for Hepatitis C Virus
LOT No.	HC006
MFG. DATE	2019-12
EXP. DATE	2021- 11
PACK SIZE	5T, 20T
STORAGE	2°C to 30°C
A.R. No. : B-HC-006a	

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.


Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 no, 20 no.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01 no.	Pass

PERFORMANCE		
CONTROLS	(Ct Values)	
Specificity Tests	HCV	Control
Negative Control	ND	29.33
Positive Control	30.13	27.8
	29.55	27.6
	29.07	28.0
	29.8	28.5
Standard deviation	± 0.44	± 0.38

\*ND- Not detected, Ct- Cycle threshold, HCV- Hepatitis C Virus

\*A deviation of  $\leq 1.5$  is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS QC APPROVED FOR SALE

  
QUALITY CONTROL  
Name: Dr. Praveen Kumar  
Designation: QC Head

Page 1 of 1

**Molbio Diagnostics Pvt. Ltd.**

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO