



MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000018

Date:23/11/2021
Doc ID: COA/TN/SL015

BRAND NAME	Truenat [®] Salmonella
GENERIC NAME	Chip-based Real Time PCR Test for Salmonella spp.
LOT No.	SL015
MFG. DATE	2021-11
EXP. DATE	2023-10
PACK SIZE	5T, 20T, 50T
STORAGE	2°C to 30°C
A.R. No. : B-SL-015	

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 dessicant pouch and 1 filter barrier tip	5 no, 20 no, 50no	Pass
2	Colour of dessicant pouch	Orange	Pass
3	Package insert	01 no.	Pass

PERFORMANCE		
CONTROLS	(Ct Values)	
Specificity Tests	SAL	Control
Negative Control	ND	29.5
Positive Control	29.14	30.33
	28.14	30.33
	28.43	29.6
	28.5	30.14
Standard deviation	± 0.42	± 0.34

*ND- Not detected, Ct- Cycle threshold, SAL- Salmonella

*A deviation of ≤ 1.5 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

for
QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head
Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO