



MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000018

Date:01/12/2021

Doc ID: COA/TN/CD135

BRAND NAME	Truenat <sup>®</sup> COVID-19
GENERIC NAME	Chip-based Real Time Duplex PCR Test for COVID-19
LOT No.	CD135
MFG. DATE	2021-10
EXP. DATE	2023-09
PACK SIZE	5T, 20T, 50T
STORAGE	2°C to 30°C
	A.R. No.: B-CD-135

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 dessiccant pouch and 1 filter barrier tip	5 no, 20 no, 50 no.	Pass
2	Colour of desiccant pouch	Orange	Pass
3	Package insert	01 no.	Pass

PERFORMANCE			
CONTROLS	(Ct Values)		
Specificity Tests	Orf1A	Egene	Control
Negative Control	ND	ND	27.13
	ND	ND	26.57
Positive Control	26.14	28	27.4
	25.6	27.17	27.43
	26.67	28	27.17
	25.78	27.67	27
	26.33	28	27.75
	26.29	27.25	27.14
	26.1	28.13	27
	26.4	27.6	27.4
Standard deviation	±0.34	±0.37	±0.26

\*ND- Not detected, Ct- Cycle threshold

\*A deviation of  $\leq 1.5$  is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS QC APPROVED FOR SALE

QUALITY CONTROL  
Name: Dr. Praveen Kumar  
Designation: QC Head

Page 1 of 1

**Molbio Diagnostics Private Limited**

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiagnostics.com

Email: sales@molbiagnostics.com Website: www.molbiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO