



MOLBIO DIAGNOSTICS PRIVATE LIMITED
ISO 13485:2016 certified company
CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 18/12/2021

Doc ID: COA/TP/VUCPK230-PLM

Product: Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit		
Lot No.: VUCPK230-PLM	Mfg. Date: 2021-10	Exp. Date: 2023-09
Pack size: 50T, 5T	Storage: 2°C to 40°C	A.R. No.: B-VUCPK-230-PLM

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab® micro PCR Analyzers			
Sr. No.	Sample Negative/Positive	Negative: Test- ND	IC- Detected
		Positive: Test- Detected	IC- Detected
		Test (Ct)	IC (Ct)
1	Negative Dengue	ND	29.86
2	Negative HBV	ND	30.4
3	Negative MTB	ND	29.33
4	Negative COVID-19	ND	26.0
5	Negative COVID-19	ND	25.75
6	Positive HBV	30.11	29.2

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO



MOLBIO DIAGNOSTICS PRIVATE LIMITED
ISO 13485:2016 certified company
CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 18/12/2021

Doc ID: COATP/VUCPK230-PLM

Product: Trueprep [®] AUTO v2 Universal Cartridge Based Sample Prep Kit		
Lot No.: VUCPK230-PLM	Mfg. Date: 2021-10	Exp. Date: 2023-09
Pack size: 50T, 5T	Storage: 2°C to 40°C	A.R. No.: B-VUCPK-230-PLM

Performance on Truelab[®] micro PCR Analyzers

Sr. No.	Sample Negative/Positive	Negative: Test- ND	IC- Detected
		Positive: Test- Detected	IC- Detected
		Test (Ct)	IC (Ct)
7	Positive HCV	27.7	29.33
8	Positive HIV	28.0	26.75
9	Positive MTB	28.17	29.33
10	Positive MTB	27.2	30.4

*Ct- Cycle threshold

*IC- Internal Control

*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE.



QUALITY CONTROL

Name: Dr. Praveen Kumar

Designation: QC Head

Page 2 of 2

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO