

MOLBIO DIAGNOSTICS PRIVATE LIMITED ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.:MFG/IVD/2019/000003

Date: 01/01/2022

Doc.ID: COA/TP/USPT232

Product: Trueprep® AUTO Univer	t: Trueprep [®] AUTO Universal Sample Pre-treatment Pack				
Lot No.: USPT232	Mfg. Date: 2021-12	Exp. Date: 2023-11			
Pack size: 50T, 20T, 5T	Storage: 2°C to 40°C	A.R. No.: B-USPT-232			

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
Α	Contents in the carton (Trueprep® AUTO Universal Sample Pre-treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Package Insert	Present	Present	
3	Disposable Transfer Pipette (1ml)	Present	Present	

	Performa	ance on	Truelab [®] micro PCR Analyzers		
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negati	ve Control				
1	Negative HBV	ND	Control should be detected	ND	29.5
2	Negative COVID-19	ND	Control should be detected	ND	25.8
3	Negative HIV	ND	Control should be detected	ND	28.75
Positiv	e Control				
1	Positive HBV	Test and Control should be detected		28.5	30.07
2	Positive HBV	Test and Control should be detected		29.0	28.5
3	Positive HCV	Test a	nd Control should be detected	27.25	29.57
4	Positive HCV	Test a	nd Control should be detected	27.71	30.8
5	Positive HIV	Test a	nd Control should be detected	28.0	29.17

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS NO SPER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE.

QUALITY CONTROL

Name: Dr. Praveen Kumar Designation:QC Head

Page 1 of 1

Molbio Diagnostics Private Limited

^{*}IC- Internal Control

^{*}ND- Not Detected