



MOLBIO DIAGNOSTICS PRIVATE LIMITED

**ISO 13485:2016 certified company
CERTIFICATE OF ANALYSIS**

Mfg. Lic. No.: MFG/IVD/2019/000040

**Date: 06/01/2022
DOCID:COA/TN/HC036**

BRAND NAME	Truenat [®] HCV
GENERIC NAME	Chip-based Real Time PCR Test for Hepatitis C Virus
LOT No.	HC036
MFG. DATE	2021-12
EXP. DATE	2023-11
PACK SIZE	5T, 20T, 50T
STORAGE	2°C to 30°C
A.R. No. : B-HC-036	

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 dessicant pouch and 1 filter barrier tip	5 no, 20 no, 50 no.	Pass
2	Colour of desiccant pouch	Orange	Pass
3	Package insert	01 no.	Pass

PERFORMANCE		
CONTROLS	(Ct Values)	
Specificity Tests	HCV	Control
Negative Control	ND	29.33
Positive Control	29.31	28.6
	29.45	28.14
	29.5	28.75
	29.18	28.43
Standard deviation	±0.14	± 0.26

*ND- Not detected, Ct- Cycle threshold, HCV- Hepatitis C Virus

*A deviation of =<1.5 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS QC APPROVED FOR SALE

QUALITY CONTROL

Name: Dr. Praveen Kumar

Designation: QC Head

Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO