

## **CERTIFICATE OF ANALYSIS**

Mfg. Lic. No.:MFG/IVD/2019/000003

Date: 14/01/2022

Doc.ID: COA/TP/USPT237

Product: Trueprep® AUTO Universal Sample Pre-treatment Pack

Lot No.: USPT237 Mfg. Date: 2022-01 Exp. Date: 2023-12

Pack size: 50T, 20T, 5T Storage: 2°C to 40°C A.R. No.: B-USPT-237

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
A	Contents in the carton (Trueprep® AUTO Universal Sample Pre-treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Package Insert	Present	Present	
3	Disposable Transfer Pipette (1ml)	Present	Present	

	Performa	ance on	Truelab <sup>®</sup> micro PCR Analyzers		
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negati	ve Control				
1	Negative HCV	ND	Control should be detected	ND	28.6
2	Negative COVID-19	ND	Control should be detected	ND	26.0
Positiv	e Control				
1	Positive HBV	Test and Control should be detected		28.14	30.57
2	Positive HBV	Test and Control should be detected		28.43	29.5
3	Positive HCV	Test and Control should be detected		30.33	30.14
4	Positive HIV	Test and Control should be detected		30.08	29.67
5	Positive HIV	Test a	nd Control should be detected	29.17	29.33
6	Positive COVID-19	Test and Control should be detected 30.17 23.6			

<sup>\*</sup>Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE.

QUALITY CONTROL

Name: Dr. Praveen Kumar

Designation: QC Head

Page 1 of 1

<sup>\*</sup>IC- Internal Control

<sup>\*</sup>ND- Not Detected