



MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2021/000071

Date:09/02/2022

Doc ID: COA/TN/004CD

BRAND NAME	Truenat <sup>®</sup> COVID-19
GENERIC NAME	Chip-based Real Time Duplex PCR Test for COVID-19
LOT No.	004CD
MFG. DATE	2022-01
EXP. DATE	2023-12
PACK SIZE	5T, 20T, 50T
STORAGE	2°C to 30°C
A.R.No.: B-004-CD	

**PRECAUTION:** Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 dessicant pouch and 1 filter barrier tip	5 no, 20 no, 50 no.	Pass
2	Colour of desiccant pouch	Orange	Pass
3	Package insert	01 no.	Pass

PERFORMANCE			
CONTROLS	(Ct Values)		
Specificity Tests	Orf 1A	E gene	Control
Negative Control	ND	ND	27.6
	ND	ND	28
Positive Control	30	29.14	29.67
	29.5	28.8	30
	30	29.13	30
	30.2	29.14	30
	30	29.13	29.4
	29.14	29	30.33
	29	28.17	29.8
	29.09	29.4	30.5
Standard deviation	± 0.48	± 0.37	± 0.34

\*ND- Not detected, Ct- Cycle threshold

\*A deviation of  $\leq 1.5$  is acceptable for Ct values

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS QC APPROVED FOR SALE PURPOSE**

*Praveen*  
  
**QUALITY CONTROL**  
For Name: Dr. Praveen Kumar  
Designation: QC Head  
Page 1 of 1

## Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO