



MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2021/000071

Date:05/05/2022

Doc ID: COA/TN/002TX

BRAND NAME	Truenat <sup>®</sup> MTB-RIF Dx
GENERIC NAME	Chip-based Real Time Duplex PCR Test for Rifampicin resistant <i>Mycobacterium tuberculosis</i>
LOT No.	002TX
MFG. DATE	2022-05
EXP. DATE	2024-04
PACK SIZE	5T, 20T, 25T, 50T
STORAGE	2°C to 30°C
	A.R. No.: B-002-TX


PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 dessicant pouch and 1 filter barrier tip	5 no, 20 no, 25 no, 50 no.	Pass
2	Colour of desiccant pouch	Orange	Pass
3	Package insert	01 no.	Pass

PERFORMANCE				
Specificity Tests	Tm		Control	Result
	P2	P1		
Sample 1	66.1	66.24	70.22	RIF RESISTANCE DETECTED
	66.98	66.43	69.96	RIF RESISTANCE DETECTED
Sample 2	72.74	66.42	70.57	RIF RESISTANCE NOT DETECTED
	72.75	66.89	70.38	RIF RESISTANCE NOT DETECTED
	72.1	66.33	71.08	RIF RESISTANCE NOT DETECTED

\*Tm-Melting temperature, RIF-Rifampicin

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS QC APPROVED FOR TESTING PURPOSE

  
QUALITY CONTROL  
Name: Dr. Praveen Kumar  
Designation: QC Head  
Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbioldiagnostics.com

Email: sales@molbioldiagnostics.com Website: www.molbioldiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO