



MOLBIO DIAGNOSTICS PVT. LTD.

An ISO 9001 & ISO 13485 Certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : 874/L

Date: 03-11-2018

PRODUCT : Truenat™ H1N1  
LOT. NO. : HN023  
MFG. DATE : 2018-10  
EXP. DATE : 2020-09  
PACK SIZE : 5T, 20T  
STORAGE : 2°C to 30°C

A.R. NO. : B-HN-023

**PRECAUTION:** The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 nos, 20 nos.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01no.	Pass

Performance			
Sr. No.	CONTROLS	Test Lot (Ct Values)	
Specificity Tests		HN	Control
Negative Control		ND	26
Positive Control		28.29	28.4
		28.4	28.2
		27.4	29.8
		27.8	29.75
Standard deviation		+ 0.46	+ 0.86

\*ND- Not detected

\*A deviation of  $\leq 1.5$  is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE



QUALITY CONTROL  
Name: Dr. Praveen Kumar  
Designation: QC Head

Page 1 of 1

Molbio Diagnostics Pvt. Ltd.

Regd. Office: H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA  
Works: Plot Nos. M 46-47, Phase III - B, Verna Industrial Estate, Verna - Goa 403 722. INDIA  
Tel: 91 832 6624783/6682000 Fax: 91 832 2783139 E mail: admin@molbiodiagnostics.com  
Office: No. UGO-7, Ground Floor, Ambrosia Residency, Cabesa, Alto Santacruz, Goa - 403 005. INDIA  
Tel: 91 832 2458111/2458112 Email: sales@molbiodiagnostics.com  
Website: www.molbiodiagnostics.com: CIN: U33125GA2000PTC002909