



MOLBIO DIAGNOSTICS PVT. LTD.

An ISO 9001:2015 & ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : MFG/IVD/2019/000018

Date: 07/11/2019

BRAND NAME : Truenat™ MTB-RIF Dx

GENERIC NAME: Chip-based Real Time PCR test for Rifampicin resistant *Mycobacterium tuberculosis*

LOT. NO. : TX036

MFG. DATE : 2019- 09

EXP. DATE : 2021- 08

PACK SIZE : 5T, 20T

STORAGE : 2°C to 30°C

A.R. NO. : B-TX-036g


PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 no, 20 no.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01 no.	Pass

PERFORMANCE				
Specificity Tests	Tm		Control	Result
	P1	P2		
Sample 1	66.08	65.46	71.72	RIF RESISTANCE DETECTED
	65.34	64.78	71.42	RIF RESISTANCE DETECTED
Sample 2	66.22	71.87	70.17	RIF RESISTANCE NOT DETECTED
	66.2	72.55	71.14	RIF RESISTANCE NOT DETECTED
	66.45	72.4	70.31	RIF RESISTANCE NOT DETECTED

*Tm-Melting temperature, RIF-Rifampicin

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO