



**MOLBIO DIAGNOSTICS PVT. LTD.**

**An ISO 9001 & ISO 13485 certified company**

**CERTIFICATE OF ANALYSIS**

Mfg. Lic. No. : 874/L

Date: 11-08-2018

**PRODUCT : Truenat™ Rabies**  
**LOT. NO. : RB006**  
**MFG. DATE : 2018- 08**  
**EXP. DATE : 2020- 07**  
**PACK SIZE : 05T, 20T**  
**STORAGE : 2°C to 30°C**

**A.R. NO. : B-RB-006**

**PRECAUTION:** The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contain 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	05 nos, 20 nos.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01 no.	Pass

Performance			
Sr. No.	CONTROLS	Test Lot (Ct Values)	
		RAB	IC
Specificity Tests			
Negative Control		ND	28
Positive Control		29.2	27
		29.14	28
		30	27
		29.17	28
Standard deviation		± 0.42	± 0.58

\*ND- Not detected

\*A deviation of  $\leq 1.5$  is acceptable for Ct values

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**



**QUALITY CONTROL**  
**Name: Dr. Praveen Kumar**  
**Designation: QC Head**

Page 1 of 1

**Molbio Diagnostics Pvt. Ltd.**

**Regd. Office:** H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA  
**Works:** Plot Nos. L 46, Phase II - D, Verna Industrial Estate, Verna - Goa 403 722 INDIA  
**Tel:** 91 832 2783267 **E mail:** admin@molbiodiagnostics.com  
**Email:** sales@molbiodiagnostics.com **Website:** www.molbiodiagnostics.com  
**CIN:** U33125GA2000PTC002909